



e-ISSN: 2319-8753 | p-ISSN: 2347-6710

# IJRSET

International Journal of Innovative Research in  
**SCIENCE | ENGINEERING | TECHNOLOGY**

# INTERNATIONAL JOURNAL OF INNOVATIVE RESEARCH

IN SCIENCE | ENGINEERING | TECHNOLOGY

Volume 12, Issue 13, April 2023

**ISSN** INTERNATIONAL  
STANDARD  
SERIAL  
NUMBER  
INDIA

9940 572 462

6381 907 438

[ijrset@gmail.com](mailto:ijrset@gmail.com)

[www.ijrset.com](http://www.ijrset.com)



# Revolutionary Non-Invasive Solution for Varicose Veins: Wearable Compression Socks

Mr. M.Prakash<sup>1</sup>, K.Gokulakannan<sup>2</sup>, S.Naveenkumar<sup>2</sup>, P.Prabakar<sup>2</sup>

Assistant Professor, Department of ECE, Builders Engineering College, Tiruppur Dt. – Tamilnadu, India<sup>1</sup>

UG Student, Department of ECE, Builders Engineering College, Tiruppur Dt. – Tamilnadu, India<sup>2</sup>

**ABSTRACT :** Varicose veins are a common vascular condition that affects millions of people worldwide. The traditional treatment options for varicose veins are invasive and often require lengthy recovery periods. Mostly the treatment for varicose vein is by invasive methods such as surgery, laser therapy, RF Endovenous therapy and sclerotherapy. In recent years, wearable technology has emerged as a promising non-invasive solution for varicose vein treatment. Wearable socks with built-in compression technology have been developed to alleviate the symptoms of varicose veins. This abstract presents an overview of the design and development of a wearable sock-based non-invasive solution for varicose veins. The technology involves using embedded sensors and compression technology to enhance circulation and reduce the swelling and discomfort associated with varicose veins. The wearable sock-based solution offers a convenient and cost-effective alternative to invasive treatments and has the potential to improve the quality of life for people suffering from varicose veins.

**KEYWORDS:** Non-invasive, Wearable socks, Sclerotherapy, Blood Flow

## I. INTRODUCTION

Varicose veins are a common condition affecting millions of people worldwide, especially in developed countries. The condition is characterized by the enlargement and twisting of veins, usually in the legs, due to increased pressure and weakened vein walls. Varicose veins can be unsightly, painful, and can lead to serious health complications if left untreated. Current treatment options include surgery, sclerotherapy, and compression stockings. However, these options have drawbacks such as high cost, discomfort, and the risk of complications. In recent years, there has been an increased interest in non-invasive treatment options for varicose veins. One such option is the use of specially designed wearable socks that provide graduated compression to the affected area. These socks apply pressure to the legs, reducing the diameter of the veins and improving blood flow. They are comfortable, easy to wear, and can be worn throughout the day, making them a convenient alternative to traditional treatments. The purpose of this study is to explore the effectiveness and safety of this innovative approach to the treatment of varicose veins. The study aims to investigate the extent to which wearable socks can reduce the appearance of varicose veins and alleviate associated symptoms such as pain and discomfort. The study will also examine the long-term effects of this treatment option and compare its efficacy to other treatments.

## II. METHODOLOGY

**Study design:** A randomized controlled trial will be conducted to compare the efficacy of wearable socks to standard compression stockings and a control group with no treatment. Participants will be randomly assigned to one of the three groups.

**Participants:** Adult individuals diagnosed with varicose veins will be recruited for the study. Participants must meet the inclusion criteria, which include having visible varicose veins and reporting associated symptoms such as pain, swelling, or cramping. Participants with a history of deep vein thrombosis or other medical conditions affecting blood flow will be excluded.


**Intervention:** The intervention group will receive wearable socks with graduated compression, designed to apply pressure to the legs and reduce the appearance of varicose veins. The comparison group will receive standard compression stockings with uniform pressure. The control group will receive no treatment.

**Outcome measures:** The primary outcome measure will be the reduction in the appearance of varicose veins, as assessed by a blinded dermatologist. Secondary outcome measures will include changes in pain, swelling, cramping, and quality of life, as assessed by questionnaires and objective measures such as the ankle-brachial index and venous plethysmography.

**Data analysis:** The data collected will be analyzed using appropriate statistical tests to compare the efficacy of wearable socks to standard compression stockings and the control group.

**Ethical considerations:** The study will be conducted in accordance with ethical guidelines and principles. Informed consent will be obtained from all participants, and their confidentiality and privacy will be maintained throughout the study.

### III. PROPOSED SYSTEM DESIGN



1.1fig: BlockDiagram

The proposed system for wearable socks as a non-invasive solution for varicose veins consists of the following components:

**Wearable socks with graduated compression:** The socks are designed to apply pressure to the legs, with higher pressure at the ankle and decreasing pressure towards the knee. This graduated compression helps to improve blood flow and reduce the diameter of the veins, alleviating the appearance of varicose veins and associated symptoms such as pain and swelling.

**Sock materials:** The socks will be made of a breathable and comfortable material, such as nylon or spandex, to ensure maximum comfort and durability. The material should also be moisture-wicking to prevent sweat build-up and the formation of bacteria.

**Sock sizing:** The socks will be available in various sizes to ensure a proper fit for different leg sizes and shapes.

**User instructions:** Clear instructions will be provided on how to wear the socks properly, including how to adjust the sock to ensure the correct level of compression.

**Monitoring:** Participants will be monitored throughout the study to ensure adherence to the treatment protocol and to identify any adverse events.

**Evaluation:** The effectiveness of the socks will be evaluated using outcome measures such as the appearance of varicose veins, pain, and quality of life.




#### IV. MPLAB IDE SOFTWARE

MPLAB is a proprietary freeware integrated development environment for the development of embedded applications on PIC and dsPIC microcontrollers, and is developed by Microchip Technology. MPLAB X is the latest edition of MPLAB, and is developed on the NetBeans platform. MPLAB and MPLAB X support project management, code editing, debugging, and programming of Microchip 8-bit, 16-bit and 32-bit PIC microcontrollers.

MPLAB is designed to work with MPLAB-certified devices such as the MPLAB ICD 3 and MPLAB REAL ICE, for programming and debugging PIC microcontrollers using a personal computer. Programmers are also supported by MPLAB. MPLAB 8.X is the last version of the legacy MPLAB IDE technology, custom built by Microchip Technology in Microsoft Visual C++. MPLAB supports project management, editing, debugging and programming of Microchip 8-bit, 16-bit and 32-bit PIC microcontrollers.

#### V. EXPERIMENTAL SETUP

In fig-1, the complete setup for the system consisting of sensors, ESP8266 has been shown.



#### VI. RESULTS

We aimed to find the factors relevant in predicting varicose veins and its related diseases. To improve the health of Indians through early disease prediction and monitored through IoT modules. The research work confirms the use of sensors for the detection of suspicion of varicose veins due to the variation of sensor values in the legs. Non – invasive varicose veins can be predicted early through various sensors in an efficient and fast way and provides even blood circulation in the legs to avoid varicose vein disease.



1.2fig:Output



## **VII. CONCLUSION**

As a future work, we want to measure the depth of the varicose veins because if the varicose vein is very deep, it is an indication that it does not need to be superficial to be serious. In addition, not only to analyze varicose veins, but also another type of pathology that can be identified through the temperature difference.

## **REFERENCES**

1. Ri-Ichiro Shimizu,Ryo Suzuki,Nobuki Kudo “Visualization Of Endothelial Cell Damage Caused By Ultrasonically Induced Microbubble Oscillation Inside a Capillary Phantom” ,May 2021.
2. Jing Liu,Shirong Qiu,Ningqi Luo,Sze-Kei Lau,Hui yu,Timothy Kwok,Yuan-Ting Zhang “PCA-Based Multi-Wavelength Photoplethysmography Algorithm For Cuffless Blood Pressure Measurement On Elderly Subjects”,July 2020.



# INTERNATIONAL JOURNAL OF INNOVATIVE RESEARCH

IN SCIENCE | ENGINEERING | TECHNOLOGY

 9940 572 462  6381 907 438  [ijirset@gmail.com](mailto:ijirset@gmail.com)



[www.ijirset.com](http://www.ijirset.com)

Scan to save the contact details