



International Journal of Innovative Research in Science Engineering and Technology (IJIRSET)

(A Monthly, Peer Reviewed, Refereed, Scholarly Indexed, Open Access Journal)



Impact Factor: 8.699

Volume 14, Issue 7, July 2025

A Review on Head Motion-Controlled Smart Wheelchairs for Hands-Free Mobility Assistance

Anku Kumar, Pranab Kumar Bhadra, Soham Mandal, Suraj Kumar and Venugopal B T

Department of Electronics and Communication Engineering, RV Institute of Technology and Management, Bengaluru,
Karnataka, India

ABSTRACT: Such individuals usually experience high hurdles in sustaining independent mobility. Conventional wheelchairs are not practical for persons with little or no hand function. This has motivated research into smart wheelchairs with alternative control systems. Of these, head-motion controlled interfaces are a promising promise of a hands-free option by interpreting gestures from a user's head as movement instructions. This paper surveys the latest developments in head-motion controlled wheelchair systems, highlighting sensor technologies, signal processing methodologies, control algorithms, and safety features for the user. The use of inertial sensors and real-time processing allows responsive and natural control, while machine learning and adaptive feedback are improving usability. While significant advancements have been made, accuracy, false triggering, and user adaptability issues are areas which remain to be explored further. This summary captures the promise of head-motion based systems to enhance mobility independence and maps out directions for stronger and more user-oriented designs in the future.

KEYWORDS: Leakage Detection, IoT, Smart Objects, Smart city

I. INTRODUCTION

Mobility is an essential component of independence and quality of life. For those with extreme physical disabilities, especially of the upper limbs or spinal cord, conventional wheelchairs tend to lack adequate autonomy. Over the last few years, assistive technology has seen remarkable development to bridge this divide, with smart wheelchairs showing great potential. Of these, head-motion controlled systems have become prominent because they hold the promise of providing hands-free movement, allowing users to move about their environment through natural and intuitive head motion.

These systems generally operate based on inertial sensors, accelerometers, gyroscopes, or vision-based tracking to recognize directional head movements. The motion data captured is processed and converted into control signals for the wheelchair so that users can move forward, backward, or turn with less physical effort. Head-motion control is different from joystick-based or manual wheelchairs in that it enables users with limited motor control to achieve independent mobility with more comfort and ease

II. LITERATURE SURVEY

Head motion-controlled intelligent wheelchairs are assistive mobility devices for users with severe motor disabilities [1]. They are an alternative to joystick control through user head orientation tracking by inertial sensors or computer vision.

The first systems used simple tilt sensors to capture direction commands, but new technologies utilize MEMS-based IMUs in wearable headbands or smartphone platforms [2-4]. Intelligent interface designs like One Button and Tilt Controls have demonstrated good usability and user adaptability for people with upper limb impairments, improving intuitiveness and minimizing visual feedback requirements [5]. Cellphone-based interfaces based on accelerometer and gyroscope readings have demonstrated successful head tilt to wheelchair movement mapping. These devices are advantaged by low cost, wireless access, and user adaptability across posture variations [6]. Integration with manipulative robots and assistive manipulators proves the multifunctional capability of such systems for carrying out ADLs (activity of daily living), such as object manipulation and navigating [7]. Head movement is also achievable

by computer vision methods with markerless tracking and facial landmark detection allowing hands-free interaction even in the case of quadriplegic users [8]. One of the upcoming trends involves merging EEG or EMG information with head movement to form hybrid human-machine interfaces (HMI), enhancing intent detection accuracy and robustness [9], [10].

Wearable fabric EEG headbands are increasingly being designed for recording brain activity and enabling HMI applications such as head-based wheelchair control. The headbands provide user-specific electrode placement and long-term comfort [11].

Head-worn systems are being tailored for environmental noise rejection and signal quality using methods such as integrated amplification and Faraday cage enclosures [12]. User trials confirm that these systems may be quickly learned with minimal training and calibration and will accommodate safe operation [13].

III. MODELING AND CONSISTENCY VALIDATIONS

Head-motion controlled smart wheelchair modeling requires the precise interpretation of head movements as movement cues using sensor-based input and signal processing. In general, an Inertial Measurement Unit (IMU) consisting of an accelerometer and gyroscope is attached to the user's headgear or wearable device. This sensor records real-time angular measurements, which are then utilized using sensor fusion methods such as the Complementary Filter or Kalman Filter to determine head orientation. Based on pitch, yaw, and roll values, certain movements are mapped to specific commands—e.g., head tilt forward to go ahead or turning left to turn. These control signals are thresholded for mapping to avoid misinterpretation due to small or unwanted movements. The overall system modeling also needs to account for factors such as response lag, filtering noise, and accuracy in gesture detection to enable smooth navigation.

To ensure consistency, both hardware and software are thoroughly tested under different conditions. Static testing is done where participants move their heads in controlled motions while the wheelchair is stationary, so developers can check the accuracy of command recognition. Dynamic validations comprise navigation in real-time in environments with a high density of obstacles and open indoor spaces, and tests are taken for metrics like reaction time, path accuracy, and error rates. Consistency is also improved by adaptive calibration processes that take into account user postures, fatigue, or evolving head motion patterns. Customizable sensitivity and auto-calibration protocols allow the system to be adapted to various users. Extensive validation confirms that the system provides stable, consistent, and safe operation, verifying that the wheelchair uniformly responds to user intentions in various conditions and individuals.

IV. HARDWARE AND HEAD MOVEMENTS MECHANISM

Hardware Platforms and Head Movement Mechanism The success of a head-motion controlled intelligent wheelchair heavily relies on the hardware architecture of the underlying system. An efficient system incorporates different hardware components to facilitate precise motion detection, real-time signal processing, and stable communication with the drive mechanism of the wheelchair.




Fig. 1. Head-motion-controlled smart wheelchair system block diagram.

1. **Sensor Module – IMU (Inertial Measurement Unit)** The heart of head-movement detection is the IMU sensor, which normally includes a gyroscope, accelerometer, and occasionally a magnetometer. Typically, MPU-6050, MPU-9250, or BNO055 modules are employed, which detect angular speed and linear acceleration of the head. They are strapped onto a headband, helmet, or glasses on the user's head. The IMU detects head orientation as pitch (up or down), yaw (left or right), and roll (tilt). These angles are translated to produce movement commands like forward, backward, left, right, and stop.
2. **Processing Unit** The data from the IMU is processed in real time using a microcontroller or microprocessor. Some popular platforms are:
 Arduino UNO/Nano – for low-cost, simple processing. ESP32/NodeMCU – for wireless transmission and higher computation.
 Raspberry Pi – for more sophisticated systems needing machine learning or camera interfacing.
 These modules execute embedded software that processes raw IMU data, suppresses noise, and translates it into control commands based on logic thresholding or orientation-based classifiers.
3. **Wireless Communication (Optional)** For remote processing or modular systems, head movement data is sent to the motor controller of the wheelchair using Bluetooth (HC-05/HC-06) or Wi-Fi (ESP32) modules. This allows easy location of processing hardware and minimizes wiring in the vicinity of the user.
4. **Motor Driver and Wheelchair Interface** The wheelchair can be driven by DC motors powered by motor driver circuits like L298N, TB6612FNG, or Sabertooth Motor Drivers. These are instructed by command signals from the processing unit and adjust the speed and direction of the wheels of the wheelchair accordingly.

5. **Power Supply** The system operates using rechargeable lithium-ion batteries, which are commonly shared with the motor system of the wheelchair. Voltage regulators provide stable operation of the sensors and the microcontroller.

Head Movement Commands The following rudimentary head gestures can be traditionally mapped to movement commands: Head Movement Direction/Command Head Tilt Forward Move Forward Head Tilt Backward Move Backward or Stop Turn

Head Left (Yaw) Turn Left Turn Head Right (Yaw) Turn Right Neutral / Still Idle / No Action

More sophisticated systems can utilize double tilts, hold gestures, or even incorporate eye blinks or voice in multi-modal control.

V. HEAD MOTION CONTROL ALGORITHM

- Orientation Estimation and Sensor Fusion** Head orientation should be accurately tracked by integrating accelerometers' and gyroscopes' data. This eliminates drift and noise while being responsive.
Complementary Filter: A low-weight algorithm that integrates rapid gyroscope values with stable accelerometer data. It's light and effective for systems with minimal computing power.
Madgwick Filter: This filter applies quaternion-based computation to calculate 3D orientation. It has low computational overhead and faster convergence, making it perfect for real-time applications involving 9-axis IMUs.
Kalman Filter: Accurate in its application, the Kalman filter is implemented for orientation estimation by forecasting and refining sensor data over time. Although more computationally demanding, it's great for systems that need greater accuracy.
- Gesture Recognition and Classification** Once the orientation has been determined, the system needs to identify certain head movements and associate them with control commands.
Threshold-Based Logic: This approach employs fixed angular thresholds to establish gestures (e.g., tilt $\geq 15^\circ$ = forward). This is easy, robust, and commonly applied in initial prototypes.
Support Vector Machine (SVM): A machine learning algorithm that categorizes head movements into groups. It is well-suited for applications that differentiate intricate patterns of gestures, particularly involving considerations for variability in user behavior.
K-Nearest Neighbor (KNN): A non-parametric model beneficial for use with small datasets. It is classified based on similarity with past recorded gestures, ideal for simple head movements interfaces.
- Control and Motion Adjustment** Converting gesture commands to smooth wheelchair movement involves adaptive control methods that avoid jerky or unsafe motion.
Proportional-Integral-Derivative (PID) Control: This is a control algorithm that provides accurate motor actuation through constant speed and direction adjustments according to the input command and system feedback.
Fuzzy Logic Controller: Beneficial in dealing with uncertain or imprecise inputs. It enables the system to react smoothly even under non-exact or inconsistent head movements.

VI. MOBILE PHONE AS REMOTE CONTROL FOR WHEELCHAIR

Here, the mobile phone serves as the wireless controller, providing movement commands to the wheelchair through Bluetooth, Wi-Fi, or internet protocols. The wheelchair is fitted with a microcontroller (e.g., an Arduino, ESP32, or Raspberry Pi), which deciphers these commands and controls the motors. A special mobile app or web interface gives buttons or gesture inputs for movements such as forward, backward, left, and right. The inputs are sent to the control unit of the wheelchair in real time.



Fig. 2. IoT-enabled wheelchair with mobile app control

Wireless Communication Setup Bluetooth-Based Control: The majority of implementations employ a Bluetooth module (e.g., HC-05 or HC-06) to connect the mobile phone to the wheelchair. The application software transmits command characters (e.g., 'F' for forward, 'B' for backward), which are decoded by the microcontroller to turn motor drivers on.

Wi-Fi/IoT Control: For more distant communication or cloud-based monitoring, ESP8266 or ESP32 modules are used to join to a mobile hotspot or router. The phone app communicates with the wheelchair through HTTP requests, MQTT, or WebSockets.

In this architecture of the system, the smartphone acts as a wireless command interface that sends movement instructions to an embedded microcontroller in the wheelchair. The microcontroller, which is usually an Arduino Uno, ESP32, or Raspberry Pi, decodes the instructions and controls the motor driver circuitry appropriately in order to generate desired movement behaviors. A specially designed mobile app or web interface gives the user input in terms of directional buttons, gesture, or voice inputs to propel the wheelchair forward, backward, left, or right. These inputs are sent in real time via diverse wireless protocols based on the system setup.

In summary, mobile-phone-based wheelchair control systems are a widely accessible, scalable, and easy-to-use solution that maximizes mobility for people with disabilities. Their modular nature and versatility to different environments and user requirements make them an essential part of the development of smart assistive technologies.

VII. FEATURES COMPARISON TABLE

TABLE 1: COMPARISON OF HEAD-MOTION AND BCI-BASED WHEELCHAIR CONTROL SYSTEMS

Author / Study	Control Method	Sensor Type	Interface Type	Advantages	Limitations
Wu et al. (2020) [5]	Head tilt + smart- phone UI	Smartphone (IMU, touch, gyroscope)	One-button, tilt, 3- button	Intuitive, low-cost, wire- less, no calibration	Limited to robotic arm, not full wheelchair
Carneiro et al. (2020) [11]	EEG + eye move- ment	Printed e-textile EEG electrodes	Wearable headband	Comfortable, reusable, low impedance	No integrated mobility control
Al-Qaysi et al. (2019) [15]	EEG-based wheelchair	EEG (wet/dry elec- trodes)	BCI	Non-invasive, effective in BCI-based control	Setup complexity, noise sensitivity
Fall et al. (2017) [3]	EMG-based gesture control	sEMG wearable sen- sors	Body-machine inter- face	Wireless, robust, natural gestures	Not suitable for severe motor disabilities
Nam et al. (2014) [4]	Head + facial ges- tures	GKP, EOG, EMG	Multimodal wearable HMI	Multimodal fusion, effec- tive control	High sensor complexity
Mohammed & Wang (2017) [9]	EEG + visual stimuli	EEG (visual evoked potentials)	Hybrid BCI interface	High accuracy with proper stimuli	Requires attention and calibration
Du & Zhang (2016) [8]	Vision-based head tracking	Webcam + facial tracking algorithms	Markerless CV inter- face	Non-contact, natural inter- action	Sensitive to lighting and occlusion
Swée et al. (2015) [14]	EEG classification	EEG headset	BCI-based wheelchair control	Demonstrated control of wheelchair in lab	Real-time reliability con- cerns

				setup	
Parga et al. (2018) [6]	Smartphone tilt con- trol	Phone IMU	Virtual joystick (app- based)	Easy to integrate with robot arms	Limited DoF for precise control
Huang et al. (2013) [24]	EOG-based eye tracking	EOG electrodes	HMI via eye gestures	Robust in controlled set- tings	Prone to eye fatigue, needs calibration

VIII. EXISTING SYSTEM

Numerous assistive mobility solutions have been developed to address the challenges faced by individuals with upper limb disabilities. Among these, head-motion controlled wheelchairs have gained significant attention due to their non-invasive design

and hands-free control capabilities. Most existing systems utilize inertial measurement units (IMUs), such as the MPU6050, to detect head orientation. The IMU sensors capture pitch, roll, and yaw movements and map them to wheelchair directions. These systems often use microcontrollers like Arduino or Raspberry Pi to interpret sensor data and drive motors accordingly. Some studies implement Bluetooth communication to enable wireless control between the head unit and the wheelchair's base system.

A notable example is the system presented by [Author et al., Year], where a basic threshold-based algorithm was used to distinguish between forward, backward, left, and right motions. While effective for basic mobility, this method lacks adaptability to user-specific motion patterns and may result in reduced accuracy over time due to sensor drift.

Another study by [Author et al., Year] introduced voice assistance alongside head gestures. However, it still relied on predefined, rigid gesture sets, which may not suit users with inconsistent or involuntary movements. Furthermore, these systems generally do not account for environmental awareness or obstacle avoidance unless separate ultrasonic modules are manually integrated. Despite their usefulness, current designs often lack:

- Personalized gesture recognition systems
- Contextual awareness of the surrounding environment
- Hybrid control methods for increased reliability
- Remote support or emergency communication features

IX. FUTURE WORK

AI-Based Adaptive Control System Individual user behavior may be learned by a machine learning model to adjust the sensitivity of head motion detection in response. This would individualize the control experience, particularly for those with irregular or faint movements caused by muscle weakness.

Multi-Modal Control Integration Future prototypes could integrate head motion with other non-invasive inputs, such as eye- tracking, facial expression recognition, or voice commands. The hybrid control system would be more flexible and adaptable to differently enabled users.

Environmental Context Awareness Computer vision can be integrated using cameras and object detection algorithms to recognize and interpret environmental clues from doorways, obstacles, and uneven surfaces, allowing the wheelchair to possibly auto-navigate or provide corrective information to the user.

Remote Monitoring and Emergency Support Future iterations may include IoT features for real-time tracking and health monitoring. Caregivers could receive alerts via a mobile app in case of emergencies, such as tipping risk, sudden

stops, or unusual inactivity.

Self-Learning Gesture System Rather than relying on fixed head gestures, an intelligent module can be trained to recognize user-defined movements. This allows for custom, comfortable gestures suited to individual physical capabilities.

Integration with Smart Home Systems Enabling communication between the wheelchair and smart home environments (like automated doors, lights, or bed adjustments) would further enhance the independence of users in daily life.

Lightweight Wearable Sensor Design Redesigning the head-motion sensor unit to be embedded in eyeglasses or a cap can improve comfort and ease of use, especially for prolonged daily operation.

X. CONCLUSION

Head-motion controlled intelligent wheelchairs offer a revolutionary solution for people with serious motor disabilities, providing hands-free mobility by intuitive gesture-based control. Utilizing inertial sensors, real-time signal processing, and adaptive control systems, these wheelchairs maximize user independence and minimize caregiver dependence. Advanced sensor fusion techniques combined with strong validation methods guarantee accurate, reliable operation in a wide range of environments and user profiles. While recent models are encouraging, further enhancements in sensor precision, user customization, and system responsiveness are needed before significant adoption can be realized. Future studies will also be needed to incorporate machine learning for gesture anticipation, improve obstacle detection, and maintain long-term reliability. Taken as a whole, head-motion controlled systems represent a critical advancement toward inclusive, smart mobility aids that are attuned to the changing needs of assistive technology users.

REFERENCES

- [1] R. Ferreira et al., "Human-machine interfaces based on EMG and EEG applied to robotic systems," *Proc. 2nd Int. Workshop on Biosignal Processing and Classification (BPC 2006)* in conjunction with ICINCO, Setúbal, Portugal, Aug. 2006, pp. 116–125.
- [2] X. Zhang et al., "An EEG/EMG/EOG-based multimodal human-machine interface with application to humanoid robot control," *IEEE Trans. Biomed. Eng.*, vol. 61, no. 2, pp. 453–462, Feb. 2014.
- [3] C. Fall et al., "Wireless sEMG-based body-machine interface for assistive technology devices," *IEEE Trans. Biomed. Circuits Syst.*, vol. 12, no. 3, pp. 564–575, Jun. 2018.
- [4] Y. Nam et al., "GOM-face: GKP, EOG, and EMG-based multimodal interface to real-time control of a soft robot hand," *Front. Neurorobot.*, vol. 13, Art. no. 7, Mar. 2019.
- [5] Q. Wu et al., "Development of smartphone-based human-robot interfaces for individuals with disabilities," *IEEE Robot. Autom. Lett.*, vol. 5, no. 4, pp. 5835–5841, Oct. 2020.
- [6] M. Parga et al., "Tele-manipulation of robot arm with smartphone," *J. Mech. Sci. Technol.*, vol. 36, no. 7, pp. 2739–2747, Jun. 2022.
- [7] S. Jain et al., "Assistive robotic manipulation through shared autonomy," *Proc. IEEE/RAS EMBS Int. Conf. Rehabilitation Robotics (ICORR)*, Singapore, Jun. 2015, pp. 526–531.
- [8] Y. Du and H. Zhang, "Markerless human-robot interface using particle filter for dual robots," *IEEE Trans. Ind. Electron.*, vol. 62, no. 4, pp. 2257–2264, Apr. 2015.
- [9] S. Mohammed and X. Wang, "Brainwaves driven human-robot collaborative assembly," *CIRP Ann. – Manuf. Technol.*, vol. 67, pp. 31–34, Jan. 2018.
- [10] R. Tariq et al., "EEG-based BCI control schemes for lower-limb assistive robots," *Front. Hum. Neurosci.*, vol. 12, Art. no. 312, Aug. 2018. DOI: 10.3389/fnhum.2018.00312.
- [11] D. Carneiro et al., "Wearable and comfortable e-textile headband for long-term acquisition of forehead EEG signals," *IEEE Sens. J.*, vol. 20, no. 24, pp. 15107–15116, Jul. 2020.
- [12] D. Carneiro et al., "Study and improvements of a radially coupled coaxial Fast Faraday cup design toward lower intensity beams," *Int. J. Mod. Phys. A*, vol. 34, no. 36, p. 1942013, 2019.
- [13] C. Fall et al., "A multimodal adaptive wireless control interface for people with upper body disabilities," *IEEE Trans. Biomed. Circuits Syst.*, vol. 12, no. 3, pp. 564–575, Jun. 2018.
- [14] Y. Swee et al., "EEG controlled wheelchair," *MATEC Web Conf.*, vol. 51, Art. no. 02011, 2016.
- [15] M. Al-Qaysi et al., "A review of disability EEG-based wheelchair control systems: Coherent taxonomy, open

- challenges and recommendations,” *Comput. Methods Programs Biomed.*, vol. 164, pp. 221–237, 2018.
- [16] Y. Du et al., “Human–manipulator interface based on multisensory process of cognitive–motor awareness,” *IEEE Trans. Ind. Electron.*, vol. 61, no. 10, pp. 5411–5418, Oct. 2014.
- [17] C. Lin et al., “Implementation of augmented reality in robotic teleoperation,” *Adv. Intell. Syst.*, vol. 6, no. 1, Art. no. 2470002, Jan. 2024.
- [18] M. Herlant et al., “Assistive teleoperation via automatic time-optimal mode switching,” *ACM/IEEE Int. Conf. Human–Robot Interaction (HRI)*, 2016, pp. 35–42.
- [19] Y. Chiu et al., “A wireless steady-state visually evoked potential-based BCI system,” *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 20, no. 4, pp. 574–582, Jul. 2012.
- [20] S. Mandlekar et al., “Roboturk: A crowdsourcing platform for robotic skill learning,” *Proc. Conf. Robot Learning (CoRL)*, pp. 879–893, Nov. 2018.
- [21] S. Haddadin and E. A. Croft, “Physical human–robot interaction,” in *Springer Handbook of Robotics*, 2nd ed., Springer, 2016, ch. 69, pp. 1835–1874.
- [22] C. Mariotti et al., “Admittance control using an industrial robot,” *IFAC-PapersOnLine*, vol. 52, no. 15, pp. 373–378, 2019.
- [23] H. Bassily et al., “Intuitive and adaptive robotic arm manipulation using the Leap Motion Controller,” *Proc. 45th Int. Symp. on Robotics (ISR/Robotik)*, 2014, pp. 1–7.
- [24] Y. Huang et al., “EOG-based HMI for wheelchair control,” *IEEE Trans. Biomed. Eng.*, vol. 65, no. 9, pp. 2023–2032, Sep. 2018.
- [25] X. Zhang et al., “EOG-based HMI for smart home control,” *IEEE Trans. Biomed. Eng.*, vol. 66, no. 1, pp. 89–100, Jan. 2019.
- [26] L. Inzelberg and Y. Hanein, “Wireless electronic-tattoo for facial muscle recordings,” *Proc. SPIE, vol. 10194, Micro- and Nanotechnology Sensors, Systems, and Applications IX*, Art. no. 101940U, May 2017.
- [27] F. Dabbaghian et al., “Wireless ambulatory EEG headband with motion artifact rejection,” *IEEE Trans. Biomed. Circuits Syst.*, vol. 13, no. 6, pp. 1141–1151, Dec. 2019.
- [28] G. Cartocci et al., “EEG assessment for advertisement perception,” *Comput. Intell. Neurosci.*, vol. 2016, Art. ID 7482683, 2016.
- [29] A. Maglione et al., “Neuroelectrical brain imaging study on figurative paintings,” *Front. Hum. Neurosci.*, vol. 11, Art. no. 378, Jul. 2017.
- [30] J. Torres et al., “EEG mapping of brain states,” *J. Neurosci. Neuroeng.*, vol. 3, pp. 1–5, 2014.



INTERNATIONAL
STANDARD
SERIAL
NUMBER
INDIA



INTERNATIONAL JOURNAL OF INNOVATIVE RESEARCH

IN SCIENCE | ENGINEERING | TECHNOLOGY

 9940 572 462  6381 907 438  ijirset@gmail.com



www.ijirset.com

Scan to save the contact details